

Allweighs Weighing Services
Westend Grange
West End Road
Epworth
Doncaster
DN9 1LA
Tel: 01427 875 867
Web: www.allweighs.co.uk



Certificate No: CJS:

2024000536

Location:

Platform Scale

Manufacturer:

Locosc LP7510B

Serial/Plant No:

20170722020

Capacity:

3000

kg

Increment:

1

kg

Frequency:

12

months

Test Date:

26 November 2024

Next due:

November 2025

Calibration Certificate of Accuracy

Business Name:

Advanced Metal Innovations Co.

Address:

Wagon Works
Rotherham Road
Sheffield
S20 1AH

Contact Name:


Lian Kingscott

Off-centre Loading Definitive:


500

Load point:	Reading:	Deviation:	Load point:	Reading:	Deviation:
1	500	0			
2	500	0			
3	500	0			
4	500	0			


☒ Four loadcell




☐ Six loadcell



☐ Single loadcell



☐ Eight loadcell



As Found

Measurement Points:	Reading:	Deviation:
0	0	0
400	400	0
800	800	0
1000	1000	0
2000	1999	1
3000	2999	1

Repeatability

3000	2999	1
------	------	---

Notes:

Platform scale found to be operating within permitted tolerances.

Definitive

Measurement Points:	Reading:	Deviation:
0	0	0
400	400	0
800	800	0
1000	1000	0
2000	2000	0
3000	3000	0

Repeatability

3000	3000	0
------	------	---

Traceable Weights Used:

W24017

Trading
Standards
Maximum
Tolerance:

3 +/-

3 kg


Calibration method and equipment used
traceable to national standards.

QMF17 Issue 5 14/07/2020

Engineer
Name:

CJS

C. Scott



Allweighs Weighing Services
Westend Grange
West End Road
Epworth
Doncaster
DN9 1LA
Tel: 01427 875 867
Web: www.allweighs.co.uk



Certificate No: CJS:

2024000494

Location:

Weighbridge

Manufacturer:

DINI Argeo 3590ET

Serial/Plant No:

83626662

Capacity:

50000

kg

Increment:

20

kg

Frequency:

12

months

Test Date:

26 November 2024

Next due:

November 2025

Calibration Certificate of Accuracy

Business Name:

Advanced Metal Innovations Co.

Address:

Wagon Works
Rotherham Road
Sheffield
S20 1AH

Contact Name:

Lian Kingscott

Off-centre Loading Definitive:



10000

Load point: Reading: Deviation: Load point: Reading: Deviation:

1	10000	0	5	10000	0
2	10000	0	6	10000	0
3	10000	0			
4	10000	0			



☐ Four loadcell

☒ Six loadcell



☐ Single loadcell

☐ Eight loadcell



As Found

Measurement Points: Reading: Deviation:

0	0	0
8000	8000	0
16000	16000	0
23700	23700	0
33700	33680	20
44160	44140	20

Repeatability

44160	44140	20
-------	-------	----

Notes:

Weighbridge found to be operating within permitted tolerances.

Definitive

Measurement Points: Reading: Deviation:

0	0	0
8000	8000	0
16000	16000	0
23700	22300	0
33700	32300	0
44160	44160	0

Repeatability

44160	44160	0
-------	-------	---

Traceable Weights Used:

W24017

Trading
Standards
Maximum
Tolerance:

60

+/-

60

kg

Calibration method and equipment used
traceable to national standards.

QMF17 Issue 5 14/07/2020

Engineer
Name:

CJS

C. Scott